




September 7, 2016

Addendum 1 to Sealed Bid 002060 Surgical Microscope Due 9/16/16 2:30PM

The specification that reads "Minimum contrast of 0.5UM" is changed to read "Minimum contrast of 0.02UM (2 percent)".


William Tremblay,
Procurement Specialist IV