



**Health Care Services Division**  
**Lallie Kemp Regional Medical Center**

June 16, 2020

**LALLIE KEMP MEDICAL CENTER**  
**ADDENDUM #2**  
**SOLICITATION # 000529**

1. Who built/manufactured the coach?
  - a. SIEMENS/KOHLER
2. Approximate age of coach?
  - a. 2006
3. Is the coach intended to remain in a fixed location and, if so, for how long
  - a. Yes with no plans to move
4. Please provide the make and model of the UPS system. What is the kvA?
  - a. 100
5. How many battery strings and how many batteries on each string. What is the mfg and model of the batteries?
  - a. Please see attached
6. Is the contractor responsible for replacing the failed batteries or battery string?
  - a. Yes
7. Please provide make and model of the Generator. What is watt size and is it electric, gas or diesel?
  - a. See attached
8. Where is the generator located?
  - a. Contained in front of unit
9. When was the last Minor and Major PM performed on the generator?
  - a. January 2020
10. If there is a power outage that takes out the chiller, and through that the compressor/coldhead become damaged, are we then responsible in the event of a quench?
  - a. YES
11. What has been the helium levels over the past 12 months as recorded by the past service provider?
  - a. SEE ATTACHED
12. Will there be an opportunity to perform a pre-inspection to identify pre-existing issues?
  - a. YES, contact Travis Thigpen @ 985-878-1297 or email [tthigp@lsuhsc.edu](mailto:tthigp@lsuhsc.edu)

**WARNING**

1. QUALIFIED PERSONS ONLY.
2. BEFORE PERFORMING MAINTENANCE, REFER TO MAINTENANCE MANUAL.
3. TO REDUCE THE RISK OF FIRE OR ELECTRIC SHOCK, INSTALL IN A TEMPERATURE AND HUMIDITY CONTROLLED INDOOR AREA FREE OF CONDUCTIVE CONTAMINANTS.

**CAUTION**

1. RISK OF ELECTRIC SHOCK.  
THIS UPS RECEIVES POWER FROM MORE THAN ONE SOURCE-DISCONNECTION OF THE AC SOURCES AND THE DC SOURCE IS REQUIRED TO DE-ENERGIZE THIS UNIT BEFORE SERVICING.
2. KEEP FILTERS CLEAN.



7 x 24 SERVICE TOLL FREE NUMBER

**(800) 887-7830**

UPS MODEL # 9800AD

UPS SERIAL # 05-7M71087660001-02

kVA 100 START UP 6-21-06

CONTRACT # MC- \_\_\_\_\_ SITE ID# \_\_\_\_\_

# KOHLER®

KOHLER CO. KOHLER WI. USA

MODEL 60REOZJB

SERIAL 2067804

SPEC. ES-67207-GA1

MFG DT: 10/05

SERVICE DUTY: STANDBY

HZ: 60 RPM: 1800 FUEL: DIESEL

		KW		KVA	
<b>SINGLE PHASE</b>					
<b>THREE PHASE</b>					
<b>VOLTAGE</b> (CODE)		120/240 (61) 1p	139/240 (61)	127/220 (61)	
<b>AMPS</b>					
277/480 (71)	220/380 (71)	120/208 (81)	347/600 (91)	110/220 (01)	
93.00					
110/220 (61) 1p	230/400 (71)	240/416 (71)			

GEN. MODEL: 4P10W  
INSUL: NEMA CLASS

BATT. 12 V  
H 40° Amb

MATL:  
60REOZJB

