

July 29, 2020

ADDENDUM NO. 1

Your reference is invited to SB # 7608 which is scheduled to open 2:00 p.m.. Thursday August 20, 2020 at Northwestern State University, Natchitoches, Louisiana.

PLEASE MAKE THE FOLLOWING CHANGES PART OF THE OFFICIAL BID:

Questions regarding Bid Specs.

On the second item, it says “Cressington 108 Auto Sputter Coater without Pump”. This type of device requires a specialized vacuum pump, but it is not listed to quote that. Is NSU planning to provide the pump itself? **No. See corrected Bid form with Pump included. The Vacuum pump shall be included in the bid price.**

If so, can you confirm the model of the pump already on hand? **NSU has provided the below corrected bid form with the model number.**

Correction to Bid Form:

See attached replacement bid form.

This addendum is hereby officially made a part of the referenced solicitation and should be attached to the bidder’s proposal or otherwise acknowledged therein.

If you have already submitted your proposal and this addendum causes you to revise your original bid, please indicate changes below and return to Business Affairs/Purchasing Section prior to bid opening in an envelope marked with the bid number, bid opening date, and time. If this addendum does not cause you to revise your bid, please acknowledge receipt of the addendum by signing your name and company below and returning it in accordance with the provisions above.

Company Name

Date

Authorized Signature

Dale Martin
Director of Purchasing

<u>Quantity</u>	<u>Model Number</u>	<u>Make</u>	<u>Alternate Model/Make</u>	<u>Description/Specifications</u>	<u>Unit Price</u>	<u>Extended Price</u>
<u>1</u>	<u>SNE-4500M Plus</u>	<u>SEC CO LTD.</u>		<p>SNE-4500M Plus Tabletop SEM Premium High-Resolution Tabletop SEM (Scanning Electron Microscope). Shall include 1-year Warranty instrument diagnostic, repair, testing and certification to manufacturer standards. All software updates shall be included with a minimum of 8-hours remote training.</p> <p>https://www.nanoimages.com/tabletop-sem-products/sne-4500m-plus-tabletop-sem/</p>	\$	\$
<u>1</u>	#7002	Cressington		<p>Cressington 108 Auto Sputter Coater with Pump. The following shall be included in bid price:</p> <p>#92080 RVP 100-3.5Dual Stage Rotary Vane Vacuum Pump</p> <p>#92080-108 All-Metal Vacuum Pump Connection Kit for 108 Series Coaters <i>Includes all vacuum connection parts to connect the Cressington Coater to the vacuum pump</i></p> <p>#7050 108 Rotary Tilting Stage for the Cressington 108 Auto Sputter Coater</p> <p>#7050-100 Crystal Head Brack Assembly (for #7050 with MTM-20)</p> <p>#9631 Universal AC Adapter for R-T Stage, 110-240V, 50/60Hz</p> <p>#93006 MTM-20 High Resolution Thickness Controller for use with Cressington 108 Auto Sputter Coater w/7050 R-T <i>Installation Included</i> <i>Includes Vacuum feedthrough, crystal holder, oscillator box, connection cables, test crystal, instructions</i> MTM-20 SPECIFICATIONS: <i>General: microprocessor based 4 digit display, push button zero</i> <i>6MHz crystal with lifetime check, 5 times per second update rate</i> <i>Thickness display range: 0-999.9nm (positive/negative)</i> <i>Resolution: 0.1nm</i> <i>Density range: 0.50 - 30.00 gm/cm3</i> <i>Crystal: 2.5mm (1/2") dia.</i> <i>Tooling factor range: 0.25 – 8.0</i> <i>Terminating facility: 0-999.9nm</i></p> <p>#93009 Thickness Monitor Crystals, pk/10</p> <p>#891-31 PELCO®Ultra 19 Rotary Pump Oil, 1 liter, 2 each.</p>	\$	\$

<u>Quantity</u>	<u>Model Number</u>	<u>Make</u>	<u>Alternate Model/Make</u>	<u>Description/Specifications</u>	<u>Unit Price</u>	<u>Extended Price</u>
				#92061 Replacement Exhaust Filter, 2 each #7032 Argon Pressure Regulator <i>Dual stage kit argon pressure regulator, working pressure gauge indicates 0-30 psi. 10' polyethylene tubing and hose clips to fit pressure regulator and Cressington Sputter Coater. Must fit standard argon gas cylinder. Note: Working pressure for sputter applications is 5-6 psi</i> #896 PELCO Bell Jar Kleen™, 14 oz. #9611 Air Filter for venting inlet to avoid particles entering the coater, complete with 1/4" connecting hose. https://www.cressington.com/product_108.html		
				<u>BID TOTAL</u>		\$

All Bids must include shipping, installation and set-up in the prices of the equipment***F.O.B:** Bid should be FOB Destination/Agency, title passing upon acceptance of merchandise. Failure to comply with this requirement may disqualify your bid.

Acknowledges receipt of the following:

ADDENDA: No: _____ Dated: _____ No: _____ Dated: _____
 No: _____ Dated: _____ No: _____ Dated: _____

 VENDOR NAME

 ADDRESS

 CITY, STATE, ZIP

 SIGNATURE AUTHORITY (Re: L.R.S. 39:1594(Act 121))

 TITLE

 FEDERAL TAX IDENTIFICATION NUMBER (FIN)

 DATE

(Members of firm or person authorized to sign bids for corporation)

BIDDERS MUST SIGN IN INK

IMPORTANT

Signature Authority: In Accordance with L.R.S.39:1594 (Act 121), the person signing the bid must be:

1. A current corporate officer, partnership member or other individual specifically authorized to submit a bid as reflected in the appropriate records on file with the Secretary of State; or
2. An individual authorized to bind the vendor as reelected by an accompanying corporate resolution, certificate or affidavit; or
3. An individual listed on the State of Louisiana Bidder's Application as authorized to execute bids.

By signing the bid, the bidder certified compliance with the above.

WE ARE AN EQUAL OPPORTUNITY UNIVERSITY