



Nicholls State University

**Purchasing Office**

P. O. Box 2052 – Thibodaux, LA 70310  
985.448.4038 – Fax: 985.448.4921

**ADDENDUM NO 2  
FOR  
NICHOLLS STATE UNIVERSITY  
THIBODAUX, LOUISIANA  
09/19/2022**

**PAGE 1 OF 1**

**Bid SB01859 – Procurement of Liquid Chromatograph Tripe-Quad Mass Spectrometer system**

**Sealed Bids are Due at 3:00 PM on 09/27/2022**

**Purpose of Addendum:**

**1) Answer Questions received**

- For sensitivity positive, Is signal to noise of at least 750,000 the minimum? Can the University provide a range of acceptance?

Answer: Sensitivity of 100,000 or better is acceptable

- Polarity switching, can the University do 50ms?

Answer: 50 ms is acceptable

RECEIPT OF SHOULD BE ACKNOWLEDGED ON THE BID RESPONSE FORM. FAILURE TO ACKNOWLEDGE ADDENDA MAY BE CAUSE FOR BID NOT TO BE CONSIDERED.

Terry G. Dupre  
Director of Purchasing, Property Control  
and Support Services Administration