INVITATION TO BID - Addendum 01	BID DUE DATE AND TIME
BOARD OF SUPERVISORS OF LOUISIANA STATE UNIVERSITY AND AGRICULTURAL & MECHANICAL COLLEGE	LSU 10/19/2023 11:00 AM CT
SOLICITATION RFQ-0000002059 SUPPLIER #	RETURN BID TO
SUPPLIER NAME AND ADDRESS	Isubids@lsu.edu
	BuyerAmy Hill BourgeoisBuyer PhoneBuyer Emailahill5@lsu.eduIssue Date10/04/2023
TITLE: MRI System for Rodent Studies	
Addendum 01: Notice is given to all parties that this solicitation is amended by the University as stated herein. This Addendum is hereby made an official part of this solicitation. See attached addendum.	
To Be Completed By Supplier	
1"No Bid" (sign and return this page only).	
My Company does not wish to receive future solicitations for this spend category.	
	days after receipt of order.
 If applicable, Supplier's Addendum Acknowledgement/Response: As an authorized agent/signatory of the supplier, I/we acknowledge receipt of this Addendum, and submit no alterations/clarifications to our original bid. 	
submit superseding revisions/clarifications to our original bid as written herein or attached hereto.	
General Instructions to Suppliers	
1. Sealed bids for furnishing the items and/or services specified are hereby solicited, and will be received by LSU Procurement at the "Return Bid To" address stated above, until the specified due date and time.	
2. Read the entire solicitation, including all terms, conditions and specifications.	
3. All bid information and prices must be typed or written in ink. Any corrections, erasures or other forms of alteration to unit price are to be initialed by the supplier.	
4. Bid prices are to be quoted FOB LSU/Destination and inclusive of any and all applicable shipping and handling charges unless otherwise specified in the solicitation. Any invoiced delivery charges not quoted and itemized on the LSU purchase order are subject to rejection and non-payment.	
 Payment is to be made within 30 days after later. 	receipt of properly executed invoice, or delivery and acceptance, whichever is
6. By signing this solicitation, the supplier certifies compliance with all general instructions to suppliers, terms, conditions and specifications; and further certifies that this bid is made without collusion or fraud.	
SUPPLIER NAME	MAILING ADDRESS
AUTHORIZED SIGNATURE	CITY, STATE ZIP
PRINTED NAME	PHONE #
TITLE	FAX #
E-MAIL	FEDERAL TAX ID #

RFQ-000002059

MRI System for Rodent Studies

Addendum 01 – Specification Revision

Inquiry

Would a 17cm bore system meet requirements?

Response

• Must be a pre-clinical high-field MRI instrument (7T) designed for rodents (mice and rats) with minimum 30cm bore.

is hereby revised to read as follows -

• Must be a pre-clinical high-field MRI instrument (7T) designed for rodents (mice and rats).