



STATE OF LOUISIANA

Office of State Procurement
INVITATION TO BID

**RESPONSES WILL BE
PUBLICLY OPENED AT THE
PHYSICAL ADDRESS BELOW**
09/17/2024
10:00 AM CST

Vendor No.: _____
Solicitation: 3000023408
Opening Date: 09/17/2024

Vendor Name and Address: (to be completed by Vendor)

Ship To Address:

1209 LEESVILLE AVENUE
BATON ROUGE, LA 70802

**TO SUBMIT AN ELECTRONIC ONLINE
RESPONSE CLICK THE LINK BELOW**

SUBMIT NON-ELECTRONIC RESPONSE TO:
Office of State Procurement
1201 N. Third Street, Suite 2-160
Baton Rouge, LA 70802

RFx Number: 3000023408

Version: 1

Buyer: RENEE BULLOCK

Buyer Phone: 225-342-8066

E-Mail: renee.bullock@la.gov

Scheduled Begin Date:

Scheduled End Date:

T-Number:

ONLINE BID RESPONSE LINK

https://lagoverpvendor.doa.louisiana.gov/rfx?sapsrm_boid=567620E9FB021EDF949F48F24F844DC4

QUESTIONS TO BE COMPLETED BY VENDOR	Required
1. _____ Have you reviewed all attachments to the bid invitation and answered all questions?	YES
2. _____ Have you entered pricing, or attached the pricing sheet (if applicable) to the bid response?	YES
3. _____ Have you attached / included all required files to the bid response?	YES
4. _____ Have you attached the signature page to the bid response?	YES
5. _____ Delivery will be made this number of days After Receipt of Order (ARO)	NO
6. _____ %discount for payment made within 30 days. Discounts for payment made in less than 30 days, of less than 1%, or applicable to an indefinite quantity contract will be accepted but will not be an award consideration.	NO

Name of Solicitation: *Non-Mandatory SiteVisit* PCR System-LDH

RFx text:

The Office of State Procurement has suspended in-person attendance by vendors at public bid openings for bids published by our office.

Any vendor who would like to view the opening of this bid can access the following link, at the date

VENDOR TELEPHONE NUMBER: EMAIL ADDRESS:	TITLE	DATE
Signature of Authorized Bidder	Name of Bidder (Typed or printed)	

and time of this bid opening:

<https://doa-ospla.zoom.us/j/2697438343>

This link will provide you with live audio and video access to this bid opening. The link will be live at 9:45 AM (Central Time) on the date of bid opening.

=====

This is an online bid opportunity. Bidders are invited to submit bids online via the link on this bid invitation. Online bidding instructions are attached to this bid invitation. All vendors must be registered in the LaGov system in order to: submit an online bid, have their bids tabulated by our office, and to receive automatic email notifications of bid opportunities.

To register as a vendor, access the following link:

https://lagoverpvendor.doa.louisiana.gov/irj/portal/anonymous?guest_user=self_reg

Please include a W-9 form with your bid if you are newly registered. Do not register again if you are already registered in the system.

Please print all attachments to ensure all documents related to this solicitation are reviewed prior to bidding.

All documents associated with this solicitation should be included in the bidder's submission.

Bid Documents Include:

Attachment A - Special Terms and Conditions - Pages 1-11

Attachment B - Price Sheet - Page 1

Attachment C - Online Bidding Instructions - Pages 1-10

Bid delivery instructions for the Office of State Procurement:

Refer to Page 1 in Attachment A - Special Terms and Conditions

=====

Method of Award:

It is the intent of the State to award this contract on an all-or-none basis to the overall lowest responsive, responsible bidder meeting the specifications. The State further reserves the right to reject individual line items from the award.

=====

Non-Mandatory Jobsite Visit:

Contact Phi Nguyen at phone # 225-219-5244 or email Phi.Nguyen@la.gov

See page 10 of Attachment A – Special Terms and Conditions for more information.

=====

This solicitation includes a trade-in allowance for the following equipment:

* Eight ThermoFisher 7500 Fast Dx PCR Systems

* Four ThermoFisher QSDX PCR Systems

If trade-ins are offered by the bidder, the cost of the trade-in allowance shall be itemized separate. Any charges for packaging and transportation of the outgoing equipment to the bidder will be paid by the bidder. The State reserves the right to exercise the trade-in option if it is deemed to be in the best interest of the State. If accepted by the State, the trade-in allowance will be subtracted from the total purchase cost. Trade-ins must be approved by the Louisiana Property Assistance Agency (LPAA).

For additional information regarding the trade-in equipment or to make an appointment for inspection, please contact: Phi Nguyen at Phi.Nguyen@la.gov. The equipment is located at 1209 Leesville Avenue, Baton Rouge, LA. 70802.

LINE	Description	Quantity	Unit	Unit Price	Extended Amount
1	<p>Product Category:41115800 PCR System</p> <p>Unit price must include cost to furnish, setup, and provide training.</p> <p>No pricing to be submitted on this page. Prices are to be submitted on Attachment B - Price Sheet.</p> <p>* Unit price must include PC and all required accessories for running the provided software and system.</p> <p>* No electrical work is required for installation</p> <p>* Literature should be included with bid response.</p> <p>* Brand: ThermoFisher or Equal</p> <p>* Model: QuantStudio 5 Dx Real-Time PCR System or Equal</p> <p>* SKU: A56200 or Equal</p> <p>* System to include the following: SKU: A47326 - QuantStudio 5 Dx Real-Time PCR System with laptop or Equal SKU: ZGP1SCQ - One Year Warranty with One PM/OQ or Equal SKU: A34911 - IQ/OQPQ, QS5 DX IVD, 96W Kit or Equal SKU: TRN00301 - Smartstart QS5 Certified training for six staff members or Equal SKU: A58195 - Diomni ENT IVD One Year License or Equal SKU: A58880 - Diomni Smartstart 0.5D or Equal SKU: A58881 - Diomni Install Consultation or Equal</p> <p>* Specifications: * Must appear on CDC's list of Group I platforms indicating it is best suited for broad public health applications. * Must be an in vitro diagnostic device intended to perform fluorescence-based PCR to provide detection of nucleic acid sequences in human-derived specimens. * Instrument system must include the instrument, software, user documentation, and either a desktop or laptop computer. * Instrument must include a heated lid assembly that heats the top half of the sample plates and provides an effective seal to minimize evaporation of the reaction mixture. * Instrument must support reaction volumes of 10–100 µL on the 0.2 mL 96-well block. * Instrument must be factory calibrated with FAM™, SYBR™ Green I, VIC™, ABY™, NED™, TAMRA™, JUN™, ROX™, Mustang Purple™, and Cy@5 dyes.</p>	7	EA		

Invitation to bid: 3000023408
 Open Date: 09/17/2024
 T-Number:

Bidder:

LINE	Description	Quantity	Unit	Unit Price	Extended Amount
	<p>* Instrument must be designed to complete a 40-cycle real-time PCR run using a fluorogenic 5' nuclease assay and Fast chemistry in under 30 minutes and the instrument must also run in standard ramping mode with standard chemistry.</p> <p>* Instrument must have real-time quantitative PCR installation specifications to demonstrate the ability to distinguish 5,000 template copies from 10,000 copies at greater than or equal to 99.7% confidence level.</p> <p>* Software must allow for locked, IVD use for cleared or approved assays.</p> <p>* Software must enable users to operate the instrument in either In Vitro Diagnostic (IVD) Mode or Test Development (DEV) Mode.</p> <p>* Software must include the ability to assign sample IDs and controls to individual wells.</p> <p>* Instrument purchase must include orientation session for approximately eight staff members by eLearning module or a trained technical scientist.</p> <p>* Vendor must supply all the necessary consumables to perform instrument qualification (IQ), operational qualification (OQ), and instrument performance verification (IPV).</p> <p>Specify Brand Bidding: _____</p> <p>Specify Model Bidding: _____</p> <p>Specify SKU Bidding: _____</p> <p>Specify Included System SKUs Bidding:</p> <p>_____</p> <p>_____</p> <p>_____</p> <p>_____</p> <p>_____</p> <p>Specify Bidding instrument appears on CDC's list of Group I Platforms: Yes _____ No _____</p> <p>Specify Bidding instrument is an in vitro diagnostic device intended to perform fluorescence-based PCR to provide detection of nucleic acid sequences in human-derived specimens: Yes _____ No _____</p> <p>Specify Bidding instrument system includes the instrument, software, user documentation, and either a desktop or laptop computer: Yes _____ No _____</p>				

LINE	Description	Quantity	Unit	Unit Price	Extended Amount
	<p>Specify Bidding instrument includes a heated lid assembly that heats the top half of the sample plates and provides an effective seal to minimize evaporation of the reaction mixture: Yes_____ No_____</p> <p>Specify Budding instrument supports reaction volumes of 10–100 µL on the 0.2 mL 96-well block: Yes_____ No_____</p> <p>Specify Bidding instrument is factory calibrated with FAM™, SYBR™ Green I, VIC™, ABY™, NED™, TAMRA™, JUN™, ROX™, Mustang Purple™, and Cy®5 dyes: Yes_____ No_____</p> <p>Specify Bidding instrument designed to complete a 40-cycle real-time PCR run using a fluorogenic 5′ nuclease assay and Fast chemistry in under 30 minutes and the instrument must also run in standard ramping mode with standard chemistry: Yes_____ No_____</p> <p>Specify Bidding instrument has real-time quantitative PCR installation specifications to demonstrate the ability to distinguish 5,000 template copies from 10,000 copies at greater than or equal to 99.7% confidence level: Yes_____ No_____</p> <p>Specify Bidding software allows for locked, IVD use for cleared or approved assays: Yes_____ No_____</p> <p>Specify Bidding software enables users to operate the instrument in either In Vitro Diagnostic (IVD) Mode or Test Development (DEV) Mode: Yes_____ No_____</p> <p>Specify Bidding software includes the ability to assign sample IDs and controls to individual wells: Yes_____ No_____</p> <p>Specify Bidding instrument purchase includes orientation session by eLearning module or a trained technical scientist: Yes_____ No_____</p> <p>Specify number of staff members for orientation session Bidding:_____</p> <p>Specify Bidding all the necessary consumables to perform instrument qualification (IQ), operational qualification (OQ), and instrument performance verification (IPV): Yes_____ No_____</p> <p>Specify Delivery Days ARO Bidding:_____</p>				

LINE	Description	Quantity	Unit	Unit Price	Extended Amount
2	<p>Product Category:41115800 PCR System</p> <p>No pricing to be submitted on this page. Prices are to be submitted on Attachment B - Price Sheet. * Unit price must include PC and all required accessories for running the provided software and system.</p> <p>* No electrical work is required for installation</p> <p>* Brand: ThermoFisher or Equal * Model: QuantStudio 7 Pro Dx Real-Time PCR System or Equal * SKU: A56199 or Equal</p> <p>* System to include the following: SKU: A51188 - QuantStudio 7 Pro Dx Real-Time PCR System or Equal SKU: A49006 - Laptop or Equal SKU: ZGP1SCQ - One Year Warranty with One PM/OQ or Equal SKU: TRN00312 - Smartstart QS7 Certified training for six staff members or Equal SKU: A58195 - Diomni ENT IVD One Year License or Equal SKU: A51464 - IQ/OQPQ or Equal SKU: A58881 - Diomni Install Consultation or Equal</p> <p>* Specifications: * Must appear on CDC's list of Group II platforms indicating it is best suited for specific public health applications. * Must be an in vitro diagnostic device intended to perform fluorescence-based PCR to provide detection of nucleic acid sequences in human-derived specimens. * Instrument system must include the instrument, software, user documentation, and if required for use, either a desktop or laptop computer. * Instrument must include a heated lid assembly that heats the top half of the sample plates and provides an effective seal to minimize evaporation of the reaction mixture. * Instrument must support reaction volumes of 10–100 µL on the 0.2 mL 96-well block. * Instrument must be factory calibrated with FAM™, SYBR™ Green I, VIC™, ABY™, NED™, TAMRA™, JUN™, ROX™, Mustang Purple™, and Cy@5 dyes. * Instrument must be designed to complete a 40-cycle real-time PCR run using an Applied Biosystems™ TaqCheck™ fast fluorogenic 5' nuclease assay and Fast</p>	5	EA		

Invitation to bid: 3000023408
 Open Date: 09/17/2024
 T-Number:

Bidder:

LINE	Description	Quantity	Unit	Unit Price	Extended Amount
	<p>chemistry in under 35 minutes and the instrument must also run in standard ramping mode with standard chemistry.</p> <p>* Instrument must have real-time quantitative PCR installation specifications to demonstrate the ability to distinguish 5,000 template copies from 10,000 copies at greater than or equal to 99.7% confidence level.</p> <p>* Software must allow locked IVD use for approved or cleared assay definition files (ADFs). ADFs would be provided by the test manufacturer.</p> <p>* Software must include the ability to assign sample IDs and controls to individual wells.</p> <p>* Software may be accessed on the device via touchscreen, when a separate computer is not required.</p> <p>* Software can perform plate QC.</p> <p>* Software must allow users to view amplification curves and threshold cycle (Ct) values.</p> <p>* Software can send results to an authorized location to support downstream reporting or LIS integration.</p> <p>* Instrument purchase must include orientation session for approximately eight staff members by eLearning module or a trained technical scientist.</p> <p>* Vendor must supply all the necessary consumables to perform instrument qualification (IQ), operational qualification (OQ), and instrument performance verification (IPV).</p> <p>Specify Brand Bidding: _____</p> <p>Specify Model Bidding: _____</p> <p>Specify SKU Bidding: _____</p> <p>Specify Included System SKUs Bidding:</p> <p>_____</p> <p>_____</p> <p>_____</p> <p>_____</p> <p>_____</p> <p>Specify Bidding instrument appears on CDC's list of Group II Platforms: Yes _____ No _____</p> <p>Specify Bidding instrument is an in vitro diagnostic device intended to perform fluorescence-based PCR to provide detection of nucleic acid sequences in human-derived specimens: Yes _____ No _____</p> <p>Specify Bidding instrument system includes the</p>				

Invitation to bid: 3000023408
 Open Date: 09/17/2024
 T-Number:

Bidder:

LINE	Description	Quantity	Unit	Unit Price	Extended Amount
	<p>instrument, software, user documentation, and either a desktop or laptop computer: Yes_____ No_____</p> <p>Specify Bidding instrument includes a heated lid assembly that heats the top half of the sample plates and provides an effective seal to minimize evaporation of the reaction mixture: Yes_____ No_____</p> <p>Specify Budding instrument supports reaction volumes of 10–100 µL on the 0.2 mL 96-well block: Yes_____ No_____</p> <p>Specify Bidding instrument is factory calibrated with FAM™, SYBR™ Green I, VIC™, ABY™, NED™, TAMRA™, JUN™, ROX™, Mustang Purple™, and Cy®5 dyes: Yes_____ No_____</p> <p>Specify Bidding instrument designed to complete a 40-cycle real-time PCR run using a Applied Biosystems™ TaqCheck™ fast fluorogenic 5′ nuclease assay and Fast chemistry in under 35 minutes: Yes_____ No_____</p> <p>Specify Bidding instrument runs in standard ramping mode with standard chemistry: Yes_____ No_____</p> <p>Specify Bidding instrument has real-time quantitative PCR installation specifications to demonstrate the ability to distinguish 5,000 template copies from 10,000 copies at greater than or equal to 99.7% confidence level: Yes_____ No_____</p> <p>Specificy Bidding software allows locked IVD use for approved or cleared assay definition files (ADFs). ADFs would be provided by the test manufacturer: Yes_____ No_____</p> <p>Specify Bidding software includes the ability to assign sample IDs and controls to individual wells: Yes_____ No_____</p> <p>Specify Bidding software may be accessed on the device via touchscreen, when a separate computer is not required: Yes_____ No_____</p> <p>Specify software can perform plate QC: Yes_____ No_____</p> <p>Specify Bidding software allows users to view amplification curves and threshold cycle (Ct) values: Yes_____ No_____</p>				

LINE	Description	Quantity	Unit	Unit Price	Extended Amount
	Specify Bidding software can send results to an authorized location to support downstream reporting or LIS integration.: Yes_____ No_____ Specify Bidding instrument purchase includes orientation session by eLearning module or a trained technical scientist: Yes_____ No_____ Specify number of staff members for orientation session Bidding:_____ Specify Bidding all the necessary consumables to perform instrument qualification (IQ), operational qualification (OQ), and instrument performance verification (IPV): Yes_____ No_____ Specify Delivery Days ARO Bidding:_____				
3	Product Category:41115800 Trade-In Allowance * No pricing to be submitted on this page. Prices are to be submitted on Attachment B - Price Sheet. * Brand: ThermoFisher * Model: 7500 Fast Dx PCR System * Serial Numbers: 275030112 - Approximately 15 years old 275030166 - Approximately 15 years old 275030632 - Approximately 10 years old 275030028 - Approximately 15 years old 275030005 - Approximately 15 years old 275030165 - Approximately 15 years old 287880213 - Approximately 9 years old 287880202 - Approximately 9 years old	8	EA		
4	Product Category:41115800 Trade-In Allowance * No pricing to be submitted on this page. Prices are to be submitted on Attachment B - Price Sheet. * Brand: ThermoFisher * Model: QSDX PCR System * Serial Numbers: 287880215 - Approximately 9 years old 287880208 - Approximately 9 years old	4	EA		

Invitation to bid: 3000023408 Open Date: 09/17/2024 T-Number:	Bidder:	Page 10 of 15
---	---------	---------------

LINE	Description	Quantity	Unit	Unit Price	Extended Amount
	287880202 - Approximately 9 years old 287880213 - Approximately 9 years old				

INSTRUCTIONS TO BIDDERS

1. READ THE ENTIRE BID, INCLUDING ALL TERMS AND CONDITIONS AND SPECIFICATIONS.
2. ALL BID PRICES MUST BE TYPED, WRITTEN IN INK OR SUBMITTED ONLINE VIA THE VENDOR PORTAL, WHICHEVER IS APPLICABLE. ANY CORRECTIONS, ERASURES OR OTHER FORMS OF ALTERATION TO UNIT PRICES SHOULD BE INITIALED BY THE BIDDER.
3. THIS BID IS TO BE MANUALLY SIGNED IN INK OR SUBMITTED ONLINE VIA THE VENDOR PORTAL, WHICHEVER IS APPLICABLE.
4. BID PRICES SHALL INCLUDE DELIVERY OF ALL ITEMS F.O.B. DESTINATION OR AS OTHERWISE PROVIDED. BIDS CONTAINING "PAYMENT IN ADVANCE" OR "C.O.D" REQUIREMENTS MAY BE REJECTED. PAYMENT IS TO BE MADE WITHIN 30 DAYS AFTER RECEIPT OF PROPERLY EXECUTED INVOICE OR DELIVERY, WHICHEVER IS LATER.
5. AMOUNT OF BID BOND REQUIRED: ADDRESSED IN THE SPECIAL TERMS AND CONDITIONS SECTION IF APPLICABLE.
6. AMOUNT OF PERFORMANCE BOND, IF REQUIRED: ADDRESSED IN THE SPECIAL TERMS AND CONDITIONS SECTION.
7. DESIRED DELIVERY: ADDRESSED IN THE SPECIAL TERMS AND CONDITIONS SECTION.
8. TO ASSURE CONSIDERATION OF YOUR BID, ALL BIDS AND ADDENDA SHOULD BE RETURNED IN AN ENVELOPE OR PACKAGE CLEARLY MARKED WITH THE BID OPENING DATE AND THE BID NUMBER, OR SUBMITTED ONLINE VIA THE VENDOR PORTAL, WHICHEVER IS APPLICABLE.
9. BIDS SUBMITTED ARE SUBJECT TO PROVISIONS OF THE LAWS OF THE STATE OF LOUISIANA INCLUDING BUT NOT LIMITED TO L.R.S. 39:1551-1736; PURCHASING RULES AND REGULATIONS; EXECUTIVE ORDERS; STANDARD TERMS AND CONDITIONS; SPECIAL TERMS AND CONDITIONS; AND SPECIFICATIONS LISTED IN THIS SOLICITATION.
10. IMPORTANT: BY SIGNING THE BID OR SUBMITTING THE BID ONLINE VIA THE VENDOR PORTAL (WHICHEVER IS APPLICABLE) THE BIDDER CERTIFIES COMPLIANCE WITH ALL INSTRUCTIONS TO BIDDERS, TERMS, CONDITIONS AND SPECIFICATIONS, AND FURTHER CERTIFIES THAT THIS BID IS MADE WITHOUT COLLUSION OR FRAUD. THIS BID IS TO BE MANUALLY SIGNED IN INK OR SUBMITTED ONLINE VIA THE VENDOR PORTAL, WHICHEVER IS APPLICABLE, BY A PERSON AUTHORIZED TO BIND THE VENDOR. ALL BID INFORMATION SHALL BE IN INK, TYPEWRITTEN, OR SUBMITTED ONLINE VIA THE VENDOR PORTAL, WHICHEVER IS APPLICABLE.
11. ADDRESS ALL INQUIRIES AND CORRESPONDENCE TO THE BUYER AT THE PHONE AND ADDRESS SHOWN ABOVE.
12. CONFERENCE: ADDRESSED IN THE SPECIAL TERMS AND CONDITIONS SECTION IF APPLICABLE.

13. BID FORMS

ALL WRITTEN BIDS, UNLESS OTHERWISE PROVIDED FOR, MUST BE SUBMITTED ON, AND IN ACCORDANCE WITH, FORMS PROVIDED AND PROPERLY SIGNED. BIDS SUBMITTED IN THE FOLLOWING MANNER WILL NOT BE ACCEPTED:

- A. BID CONTAINS NO SIGNATURE INDICATING INTENT TO BE BOUND;
- B. BID FILLED OUT IN PENCIL; AND
- C. BID NOT SUBMITTED ON THE STATE'S STANDARD FORMS

BIDS MUST BE RECEIVED AT THE ADDRESS SPECIFIED IN THE SOLICITATION OR SUBMITTED ONLINE VIA THE VENDOR PORTAL, WHICHEVER IS APPLICABLE, PRIOR TO BID OPENING TIME AND DATE INDICATED IN ORDER TO BE CONSIDERED. FAX ALTERATIONS, NOT INVOLVING PRICING, TO BIDS RECEIVED BEFORE BID OPENING TIME WILL BE CONSIDERED PROVIDED BID FORM AND FAXED ALTERATION HAVE BEEN RECEIVED AND TIME-STAMPED BEFORE BID OPENING. ALTERATIONS TO PRICE MUST BE RECEIVED AT THE ADDRESS LISTED IN THE SOLICITATION OR VIA THE ONLINE VENDOR PORTAL BEFORE THE BID OPENING TIME AND DATE. ENTIRE BID SHOULD BE RETURNED, EXCEPT ITEM PAGES NOT BID.

14. STANDARDS OF QUALITY

ANY PRODUCT OR SERVICE BID SHALL CONFORM TO ALL APPLICABLE FEDERAL AND STATE LAWS AND REGULATIONS AND THE SPECIFICATIONS CONTAINED IN THE SOLICITATION. UNLESS OTHERWISE SPECIFIED IN THE SOLICITATION, AND ANY

MANUFACTURER'S NAME, TRADE NAME, BRAND NAME, OR CATALOG NUMBER USED IN THE SPECIFICATION IS FOR THE PURPOSE OF DESCRIBING THE STANDARD OF QUALITY, PERFORMANCE AND CHARACTERISTICS DESIRED AND IS NOT INTENDED TO LIMIT OR RESTRICT COMPETITION. BIDDER MUST SPECIFY THE BRAND AND MODEL NUMBER OF THE PRODUCT OFFERED IN HIS BID. BIDS NOT SPECIFYING BRAND AND MODEL NUMBER SHALL BE CONSIDERED AS OFFERING THE EXACT PRODUCTS SPECIFIED IN THE SOLICITATION.

15. DESCRIPTIVE INFORMATION

BIDDERS PROPOSING AN EQUIVALENT BRAND OR MODEL SHOULD SUBMIT WITH THE BID INFORMATION (SUCH AS ILLUSTRATIONS, DESCRIPTIVE LITERATURE, TECHNICAL DATA) SUFFICIENT FOR STATE OF LOUISIANA TO EVALUATE QUALITY, SUITABILITY, AND COMPLIANCE WITH THE SPECIFICATIONS IN THE SOLICITATION. FAILURE TO SUBMIT DESCRIPTIVE INFORMATION MAY CAUSE BID TO BE REJECTED. ANY CHANGE MADE TO A MANUFACTURER'S PUBLISHED SPECIFICATIONS SUBMITTED FOR A PRODUCT SHALL BE VERIFIABLE BY THE MANUFACTURER. IF ITEM(S) BID DO NOT FULLY COMPLY WITH SPECIFICATIONS (INCLUDING BRAND AND/OR PRODUCT NUMBER), BIDDER SHOULD STATE IN WHAT RESPECT ITEM(S) DEVIATE. FAILURE TO NOTE EXCEPTIONS ON THE BID FORM WILL NOT RELIEVE THE SUCCESSFUL BIDDER(S) FROM SUPPLYING THE ACTUAL PRODUCTS REQUESTED.

16. BID OPENING

BIDDERS MAY ATTEND THE BID OPENING, BUT NO INFORMATION OR OPINIONS CONCERNING THE ULTIMATE CONTRACT AWARD WILL BE GIVEN AT THE BID OPENING OR DURING THE EVALUATION PROCESS. BIDS MAY BE EXAMINED IN ACCORDANCE WITH LA R.S. 44.I. INFORMATION PERTAINING TO COMPLETED FILES MAY BE SECURED BY VISITING THE STATE OF LOUISIANA DURING NORMAL WORKING HOURS.

17. PRICES

UNLESS OTHERWISE SPECIFIED BY THE STATE OF LOUISIANA IN THE SOLICITATION, BID PRICES MUST BE COMPLETE, INCLUDING TRANSPORTATION PREPAID BY BIDDER TO DESTINATION AND FIRM FOR ACCEPTANCE FOR A MINIMUM OF 30 DAYS. IF ACCEPTED, PRICES MUST BE FIRM FOR THE CONTRACTUAL PERIOD. BIDS OTHER THAN F.O.B. DESTINATION MAY BE REJECTED. PRICES SHOULD BE QUOTED IN THE UNIT (EACH, BOX, CASE, ETC.) AS SPECIFIED IN THE SOLICITATION.

18. DELIVERIES

BIDS MAY BE REJECTED IF THE DELIVERY TIME INDICATED IS LONGER THAN THAT SPECIFIED IN THE SOLICITATION.

19. TAXES

VENDOR IS RESPONSIBLE FOR INCLUDING ALL APPLICABLE TAXES IN THE BID PRICE. STATE AGENCIES ARE EXEMPT FROM ALL STATE AND LOCAL SALES AND USE TAXES.

20. NEW PRODUCTS

UNLESS SPECIFICALLY CALLED FOR IN THE SOLICITATION, ALL PRODUCTS FOR PURCHASE MUST BE NEW, NEVER PREVIOUSLY USED, AND THE CURRENT MODEL AND/OR PACKAGING. NO REMANUFACTURED, DEMONSTRATOR, USED OR IRREGULAR PRODUCT WILL BE CONSIDERED FOR PURCHASE UNLESS OTHERWISE SPECIFIED IN THE SOLICITATION. THE MANUFACTURER'S STANDARD WARRANTY WILL APPLY UNLESS OTHERWISE SPECIFIED IN THE SOLICITATION.

21. CONTRACT RENEWALS

UPON AGREEMENT OF THE STATE OF LOUISIANA AGENCY AND THE CONTRACTOR, A TERM CONTRACT MAY BE EXTENDED FOR 2 ADDITIONAL 12-MONTH PERIODS AT THE SAME PRICES, TERMS AND CONDITIONS. IN SUCH CASES, THE TOTAL CONTRACT TERM CANNOT EXCEED 36 MONTHS.

22. CONTRACT CANCELLATION

THE STATE OF LOUISIANA HAS THE RIGHT TO TERMINATE THE CONTRACT IMMEDIATELY FOR ANY OF THE FOLLOWING REASONS:

(A) MISREPRESENTATION BY THE CONTRACTOR; (B) CONTRACTOR'S FRAUD, COLLUSION, CONSPIRACY OR OTHER UNLAWFUL MEANS OF OBTAINING ANY CONTRACT WITH THE STATE OF LOUISIANA; (C) CONFLICT OF CONTRACT PROVISIONS WITH CONSTITUTIONAL OR STATUTORY PROVISIONS OF STATE OR FEDERAL LAW; (D) ABUSIVE OR BELLIGERENT CONDUCT BY CONTRACTOR TOWARDS AN EMPLOYEE OR AGENT OF THE STATE; (E) CONTRACTOR'S INTENTIONAL VIOLATION OF THE PROCUREMENT CODE (LA R.S. 39:1551 ET SEQ.) AND ITS CORRESPONDING REGULATIONS; OR, (F) ANY LISTED REASON FOR DEBARMENT UNDER LA R.S. 39:1672.

THE STATE OF LOUISIANA MAY TERMINATE THE CONTRACT FOR CONVENIENCE AT ANY TIME (1) BY GIVING THIRTY (30) DAYS WRITTEN NOTICE TO THE CONTRACTOR OF SUCH TERMINATION: OR (2) BY NEGOTIATING WITH THE CONTRACTOR AN EFFECTIVE DATE. THE STATE SHALL PAY CONTRACTOR FOR, IF APPLICABLE: (A) DELIVERABLES IN PROGRESS; (B) THE PERCENTAGE THAT HAS BEEN COMPLETED SATISFACTORILY; AND, (C) FOR TRANSACTION-BASED SERVICES UP TO THE DATE OF TERMINATION, TO THE EXTENT WORK HAS BEEN PERFORMED SATISFACTORILY.

THE STATE OF LOUISIANA HAS THE RIGHT TO TERMINATE THE CONTRACT FOR CAUSE BY GIVING THIRTY (30) DAYS WRITTEN NOTICE TO THE CONTRACTOR OF SUCH TERMINATION FOR ANY OF THE FOLLOWING NON-EXCLUSIVE REASONS: (A) FAILURE TO DELIVER WITHIN THE TIME SPECIFIED IN THE CONTRACT; (B) FAILURE OF THE PRODUCT OR SERVICE TO MEET SPECIFICATIONS, CONFORM TO SAMPLE QUALITY OR TO BE DELIVERED IN GOOD CONDITION; OR, (C) ANY OTHER BREACH OF CONTRACT.

23. DEFAULT OF CONTRACTOR

FAILURE TO DELIVER WITHIN THE TIME SPECIFIED OR TO PROVIDE THE SERVICES AS SPECIFIED IN THE BID WILL CONSTITUTE A DEFAULT AND MAY CAUSE CANCELLATION OF THE CONTRACT. WHERE THE STATE HAS DETERMINED THE CONTRACTOR TO BE IN DEFAULT, THE STATE RESERVES THE RIGHT TO PURCHASE ANY OR ALL PRODUCTS OR SERVICES COVERED BY THE CONTRACT ON THE OPEN MARKET AND TO CHARGE THE CONTRACTOR WITH COST IN EXCESS OF THE CONTRACT PRICE. UNTIL SUCH ASSESSED CHARGES HAVE BEEN PAID, NO SUBSEQUENT BID FROM THE DEFAULTING CONTRACTOR WILL BE CONSIDERED.

24. ORDER OF PRIORITY

IN THE EVENT THERE IS A CONFLICT BETWEEN THE INSTRUCTIONS TO BIDDERS OR STANDARD CONDITIONS AND THE SPECIAL TERMS AND CONDITIONS, THE SPECIAL TERMS AND CONDITIONS SHALL GOVERN.

25. APPLICABLE LAW

ALL CONTRACTS SHALL BE CONSTRUED IN ACCORDANCE WITH AND GOVERNED BY THE LAWS OF THE STATE OF LOUISIANA.

26. COMPLIANCE WITH CIVIL RIGHTS LAWS

BY SUBMITTING AND SIGNING THIS BID, BIDDER AGREES TO ABIDE BY THE REQUIREMENTS OF THE FOLLOWING AS APPLICABLE: TITLE VI AND VII OF THE CIVIL RIGHTS ACT OF 1964, AS AMENDED BY THE EQUAL EMPLOYMENT OPPORTUNITY ACT OF 1972, FEDERAL EXECUTIVE ORDER 11246, FEDERAL REHABILITATION ACT OF 1973, AS AMENDED, THE VIETNAM ERA VETERAN'S READJUSTMENT ASSISTANCE ACT OF 1974, TITLE IX OF THE EDUCATION AMENDMENTS OF 1972, THE AGE DISCRIMINATION ACT OF 1975, AND BIDDER AGREES TO ABIDE BY THE REQUIREMENTS OF THE AMERICANS WITH DISABILITIES ACT OF 1990. BIDDER AGREES NOT TO DISCRIMINATE IN ITS EMPLOYMENT PRACTICES, AND WILL RENDER SERVICES UNDER ANY CONTRACT ENTERED INTO AS A RESULT OF THIS SOLICITATION WITHOUT REGARD TO RACE, COLOR, RELIGION, SEX, SEXUAL ORIENTATION, NATIONAL ORIGIN, VETERAN STATUS, POLITICAL AFFILIATION, DISABILITY, OR AGE IN ANY MATTER RELATING TO EMPLOYMENT. ANY ACT OF DISCRIMINATION COMMITTED BY BIDDER, OR FAILURE TO COMPLY WITH THESE STATUTORY OBLIGATIONS WHEN APPLICABLE, SHALL BE GROUNDS FOR TERMINATION OF ANY CONTRACT ENTERED INTO AS A RESULT OF THIS SOLICITATION.

27. ADA ACCESSIBILITY REQUIREMENTS

IF APPLICABLE, THE CONTRACTOR WARRANTS IT WILL COMPLY WITH FEDERAL AND STATE DISABILITIES LAWS AND REGULATIONS AND ALSO WARRANTS THAT THE PRODUCTS AND SERVICES PROVIDED CONFORM TO THE APPLICABLE ACCESSIBILITY REQUIREMENTS OF WCAG 2.1 LEVEL AA OR THE MOST CURRENT VERSION (THE "ACCESSIBILITY STANDARDS"), SECTION 508 OF THE REHABILITATION ACT OF 1973, AND THE AMERICANS WITH DISABILITIES ACT. THE CONTRACTOR AGREES TO PROMPTLY RESPOND TO AND RESOLVE ANY COMPLAINT REGARDING ACCESSIBILITY OF ITS PRODUCTS AND SERVICES. IF AT ANY TIME, PRODUCTS AND SERVICES PROVIDED UNDER THIS CONTRACT DO NOT FULLY CONFORM TO THE ACCESSIBILITY STANDARDS, CONTRACTOR SHALL IMMEDIATELY ADVISE THE STATE OF LOUISIANA OFFICE OF STATE PROCUREMENT (THE "STATE") IN WRITING OF THE NONCONFORMANCE AND SHALL PROVIDE THE STATE A PLAN TO ACHIEVE CONFORMANCE TO THE ACCESSIBILITY STANDARDS, INCLUDING BUT NOT LIMITED TO, AN INTENDED TIMELINE FOR CONFORMANCE. CONTRACTOR FURTHER AGREES TO INDEMNIFY AND HOLD HARMLESS THE STATE FROM ANY CLAIMS OR DAMAGES ARISING OUT OF ITS FAILURE TO COMPLY WITH THE REQUIREMENTS OF THIS PARAGRAPH. FAILURE TO COMPLY WITH THESE REQUIREMENTS SHALL CONSTITUTE A MATERIAL BREACH OF THIS CONTRACT AND MAY BE GROUNDS FOR TERMINATION OF THIS CONTRACT BY THE STATE.

28. SPECIAL ACCOMMODATION

ANY "QUALIFIED INDIVIDUAL WITH A DISABILITY" AS DEFINED BY THE AMERICANS WITH DISABILITIES ACT WHO HAS SUBMITTED A BID AND DESIRES TO ATTEND THE BID OPENING, MUST NOTIFY THIS OFFICE IN WRITING NOT LATER THAN 7 DAYS PRIOR TO THE

BID OPENING DATE OF THEIR NEED FOR SPECIAL ACCOMMODATIONS. IF THE REQUEST CANNOT BE REASONABLY PROVIDED, THE INDIVIDUAL WILL BE INFORMED PRIOR TO THE BID OPENING.

29. INDEMNITY

CONTRACTOR AGREES, UPON RECEIPT OF WRITTEN NOTICE OF A CLAIM OR ACTION, TO DEFEND THE CLAIM OR ACTION, OR TAKE OTHER APPROPRIATE MEASURE, TO INDEMNIFY, AND HOLD HARMLESS, THE STATE, ITS OFFICERS, ITS AGENTS AND ITS EMPLOYEES FROM AND AGAINST ALL CLAIMS AND ACTIONS FOR BODILY INJURY, DEATH OR PROPERTY DAMAGES CAUSED BY THE FAULT OF THE CONTRACTOR, ITS OFFICERS, ITS AGENTS, OR ITS EMPLOYEES. CONTRACTOR IS OBLIGATED TO INDEMNIFY ONLY TO THE EXTENT OF THE FAULT OF THE CONTRACTOR, ITS OFFICERS, ITS AGENTS, OR ITS EMPLOYEES. HOWEVER, THE CONTRACTOR SHALL HAVE NO OBLIGATION AS SET FORTH ABOVE WITH RESPECT TO ANY CLAIM OR ACTION FROM BODILY INJURY, DEATH OR PROPERTY DAMAGES ARISING OUT OF THE FAULT OF THE STATE, ITS OFFICERS, ITS AGENTS OR ITS EMPLOYEES.

30. SIGNATURE AUTHORITY

ATTENTION: R.S. 39:1594(C)(4) REQUIRES EVIDENCE OF AUTHORITY TO SIGN AND SUBMIT BIDS TO THE STATE OF LOUISIANA. YOU MUST INDICATE WHICH OF THE FOLLOWING APPLY TO THE SIGNER OF THIS BID.

FOR THE PURPOSES OF LAGOV SUPPLIER PORTAL BIDDING, THE SIGNATURE OF AUTHORIZED SIGNATORY OF BIDDER IS REPRESENTED BY THE LAGOV SUPPLIER PORTAL USER LOG-IN ID. THE NAME OF AUTHORIZED SIGNATORY OF BIDDER IS REPRESENTED BY THE CONTACT PERSON ASSOCIATED WITH THE LAGOV SUPPLIER PORTAL USER LOG-IN ID.

PLEASE CIRCLE ONE:

1. THE SIGNER OF THIS BID IS EITHER A CORPORATE OFFICER WHO IS LISTED ON THE MOST CURRENT ANNUAL REPORT ON FILE WITH THE SECRETARY OF STATE OR A MEMBER OF A PARTNERSHIP OR PARTNERSHIP IN COMMENDAM AS REFLECTED IN THE MOST CURRENT PARTNERSHIP RECORDS ON FILE WITH THE SECRETARY OF STATE. A COPY OF THE ANNUAL REPORT OR PARTNERSHIP RECORD MUST BE SUBMITTED TO THIS OFFICE BEFORE CONTRACT AWARD.

2. THE SIGNER OF THE BID IS A REPRESENTATIVE OF THE BIDDER AUTHORIZED TO SUBMIT THIS BID AS EVIDENCED BY DOCUMENTS SUCH AS, CORPORATE RESOLUTION, CERTIFICATION AS TO CORPORATE PRINCIPAL, ETC. IF THIS APPLIES A COPY OF THE RESOLUTION, CERTIFICATION, OR OTHER SUPPORTIVE DOCUMENTS SHOULD BE ATTACHED HERETO.

3. THE BIDDER HAS FILED WITH THE SECRETARY OF STATE AN AFFIDAVIT OR RESOLUTION OR OTHER ACKNOWLEDGED/ AUTHENTIC DOCUMENT INDICATING THAT THE SIGNER IS AUTHORIZED TO SUBMIT BIDS FOR PUBLIC CONTRACTS. A COPY OF THE APPLICABLE DOCUMENT MUST BE SUBMITTED TO THIS OFFICE BEFORE CONTRACT AWARD.

4. THE SIGNER OF THE BID HAS BEEN DESIGNATED BY THE BIDDER AS AUTHORIZED TO SUBMIT BIDS ON THE BIDDER'S VENDOR REGISTRATION ON FILE WITH THIS OFFICE.

31. IN ACCORDANCE WITH THE PROVISIONS OF R.S. 39: 2192, IN AWARDING CONTRACTS AFTER AUGUST 15, 2010, ANY PUBLIC ENTITY IS AUTHORIZED TO REJECT A PROPOSAL OR BID FROM, OR NOT AWARD THE CONTRACT TO, A BUSINESS IN WHICH ANY INDIVIDUAL WITH AN OWNERSHIP INTEREST OF FIVE PERCENT OR MORE, HAS BEEN CONVICTED OF, OR HAS ENTERED A PLEA OF GUILTY OR NOLO CONTENDERE TO ANY STATE FELONY OR EQUIVALENT FEDERAL FELONY CRIME COMMITTED IN THE SOLICITATION OR EXECUTION OF A CONTRACT OR BID AWARDED UNDER THE LAWS GOVERNING PUBLIC CONTRACTS UNDER THE PROVISIONS OF CHAPTER 10 OF TITLE 38 OF THE LOUISIANA REVISED STATUTES OF 1950, OR THE LOUISIANA PROCUREMENT CODE UNDER THE PROVISIONS OF CHAPTER 17 OF TITLE 39.

32. FEDERAL CLAUSES, IF APPLICABLE

ANTI-KICKBACK CLAUSE

THE CONTRACTOR HEREBY AGREES TO ADHERE TO THE MANDATE DICTATED BY THE COPELAND "ANTI-KICKBACK" ACT WHICH PROVIDES THAT EACH CONTRACTOR OR SUBGRANTEE SHALL BE PROHIBITED FROM INDUCING, BY ANY MEANS, ANY PERSON EMPLOYED IN THE COMPLETION OF WORK, TO GIVE UP ANY PART OF THE COMPENSATION TO WHICH HE IS OTHERWISE ENTITLED.

CLEAN AIR ACT

THE CONTRACTOR HEREBY AGREES TO ADHERE TO THE PROVISIONS WHICH REQUIRE COMPLIANCE WITH ALL APPLICABLE STANDARDS, ORDERS OR REQUIREMENTS ISSUED UNDER SECTION 306 OF THE CLEAN AIR ACT WHICH PROHIBITS THE USE UNDER NON-EXEMPT CONTRACTS, GRANTS OR LOANS OF FACILITIES INCLUDED ON THE EPA LIST OF VIOLATING FACILITIES.

ENERGY POLICY AND CONSERVATION ACT

THE CONTRACTOR HEREBY RECOGNIZES THE MANDATORY STANDARDS AND POLICIES RELATING TO ENERGY EFFICIENCY WHICH ARE CONTAINED IN THE STATE ENERGY CONSERVATION PLAN ISSUED IN COMPLIANCE WITH THE ENERGY POLICY AND CONSERVATION ACT (P.L. 94-163).

CLEAN WATER ACT

THE CONTRACTOR HEREBY AGREES TO ADHERE TO THE PROVISIONS WHICH REQUIRE COMPLIANCE WITH ALL APPLICABLE STANDARDS, ORDERS OR REQUIREMENTS ISSUED UNDER SECTION 508 OF THE CLEAN WATER ACT WHICH PROHIBITS THE USE UNDER NON-EXEMPT FEDERAL CONTRACTS, GRANTS OR LOANS OF FACILITIES INCLUDED ON THE EPA LIST OF VIOLATING FACILITIES.

ANTI-LOBBYING AND DEBARMENT ACT

THE CONTRACTOR WILL BE EXPECTED TO COMPLY WITH FEDERAL STATUTES IN THE ANTI-LOBBYING ACT AND THE DEBARMENT ACT.

33. PROHIBITION OF DISCRIMINATORY BOYCOTTS OF ISRAEL

IN ACCORDANCE WITH LA R.S. 39:1602.1, THE FOLLOWING APPLIES TO ANY BID WITH A VALUE OF \$100,000 OR MORE AND TO VENDORS WITH FIVE OR MORE EMPLOYEES:

BY SUBMITTING A RESPONSE TO THIS SOLICITATION, THE BIDDER OR PROPOSER CERTIFIES AND AGREES THAT THE FOLLOWING INFORMATION IS CORRECT: IN PREPARING ITS RESPONSE, THE BIDDER OR PROPOSER HAS CONSIDERED ALL PROPOSALS SUBMITTED FROM QUALIFIED, POTENTIAL SUBCONTRACTORS AND SUPPLIERS, AND HAS NOT, IN THE SOLICITATION, SELECTION, OR COMMERCIAL TREATMENT OF ANY SUBCONTRACTOR OR SUPPLIER, REFUSED TO TRANSACT OR TERMINATED BUSINESS ACTIVITIES, OR TAKEN OTHER ACTIONS INTENDED TO LIMIT COMMERCIAL RELATIONS, WITH A PERSON OR ENTITY THAT IS ENGAGING IN COMMERCIAL TRANSACTIONS IN ISRAEL OR ISRAEL-CONTROLLED TERRITORIES, WITH THE SPECIFIC INTENT TO ACCOMPLISH A BOYCOTT OR DIVESTMENT OF ISRAEL. THE BIDDER HAS ALSO NOT RETALIATED AGAINST ANY PERSON OR OTHER ENTITY FOR REPORTING SUCH REFUSAL, TERMINATION, OR COMERCIAALLY LIMITING ACTIONS. THE STATE RESERVES THE RIGHT TO REJECT THE RESPONSE OF THE BIDDER OR PROPOSER IF THIS CERTIFICATION IS SUBSEQUENTLY DETERMINED TO BE FALSE, AND TO TERMINATE ANY CONTRACT AWARDED BASED ON SUCH A FALSE RESPONSE.

34. CERTIFICATION OF NO FEDERAL SUSPENSION OR DEBARMENT

BY SIGNING AND SUBMITTING ANY BID FOR \$25,000 OR MORE, THE BIDDER CERTIFIES THAT THEIR COMPANY, ANY SUBCONTRACTORS, OR PRINCIPALS ARE NOT SUSPENDED OR DEBARRED BY THE GENERAL SERVICES ADMINISTRATION (GSA) IN ACCORDANCE WITH THE REQUIREMENTS IN "AUDIT REQUIREMENTS IN SUBPART F OF THE OFFICE OF MANAGEMENT AND BUDGET'S UNIFORM ADMINISTRATIVE REQUIREMENTS, COST PRINCIPLES, AND AUDIT REQUIREMENTS FOR FEDERAL AWARDS" (FORMERLY OMB CIRCULAR A-133).

A LIST OF PARTIES WHO HAVE BEEN SUSPENDED OR DEBARRED CAN BE VIEWED VIA THE INTERNET AT [HTTPS://WWW.SAM.GOV](https://www.sam.gov).

35. SECRETARY OF STATE REGISTRATION

IN ACCORDANCE WITH LOUISIANA LAW, ALL CORPORATIONS (SEE, LA R.S. 12:262.1) AND LIMITED LIABILITY COMPANIES (SEE, LA R.S. 12:1308.2) MUST BE REGISTERED AND IN GOOD STANDING WITH THE LOUISIANA SECRETARY OF STATE IN ORDER TO HOLD A PURCHASE ORDER AND/OR CONTRACT OVER \$25,000.

36. ALL BID AMOUNTS SHALL BE SUBMITTED IN UNITED STATES DOLLARS.

37. INDEPENDENT PRICE DETERMINATION

BY SIGNING AND SUBMITTING A BID, THE BIDDER CERTIFIES THAT THE PRICE SUBMITTED WAS INDEPENDENTLY ARRIVED AT WITHOUT COLLUSION.

38. E-VERIFY

THE CONTRACTOR SHALL COMPLY WITH THE PROVISIONS OF LA R.S. 23:995 AND FEDERAL LAW PERTAINING TO E-VERIFY IN THE PERFORMANCE OF SERVICES UNDER THE CONTRACT.