

**OFFICE OF PURCHASING  
O'KEEFE ADMINISTRATION BUILDING**  
501 City Park Avenue, Bldg 37  
New Orleans, LA 70119  
(504) 762-3027 FAX: (504) 762.3089  
www.dcc.edu

## Addendum No: (1)

Bid No: 40006-R0028475  
Bid Name: Camera Equipment  
Opening Date: November 7, 2024 at 2:00PM CST

Addendum No: 1  
Date Issued: October 30, 2024

*This addendum is hereby officially made a part of the referenced solicitation. The below has been clarified as follows.*

1	Will the college accept a camera with all the requirements in the bid but with 8MP instead of 12MP?	No, must be a minimum of 12MP. See Page 6 for Substitute, Equivalent instructions.
2	The camera connects by the universal C-mount to Trinocular microscopes. Is the word Binoc and Bino in the specifications is a mistake?	It has both types of C-Mounts.
3	Are the microscopes equipped with Nikon's C-mount adapters?	Yes, microscopes have the C-mount adapters from Nikon.
4	Would your colleagues be open to a 13in screen rather than a 15in screen?	See Page 6 for Substitute, Equivalent instructions.

**Acknowledgement:** Acknowledgement of this addendum is mandatory.

**Bidder shall acknowledge where indicated on the Bid Form or if your Bid has already been submitted, please sign below where indicted and email this form in its entirety to [wboesc@dcc.edu](mailto:wboesc@dcc.edu), it can also be mailed to the address stated on page 2 of the Bid Document. Whether indicated on the Bid Form, by email, or mail, acknowledgement of this Addendum must be legible and received prior to the opening date and time indicated in the bid documents. **Failure to acknowledge and/or submit addendum by the bid due date and time will cause your bid to be disqualified.****

The below signed acknowledges receipt of this addendum.

Company: \_\_\_\_\_ Signature: \_\_\_\_\_

Printed Name: \_\_\_\_\_