INVITATION TO BID - Addendum 02			BID DUE DATE AND TIME		
LOUISIA	OF SUPERVISORS OF NA STATE UNIVERSITY RICULTURAL & MECHANICAL COLLEGE	LSU	02/18/2025	11:00 AM CT	
SOLICITATION RFQ-0000002436 SUPPLIER #			RETURN BID TO		
SUPP	LIER NAME AND ADDRESS			_	
			Isubids@lst	u.edu	
			Buyer Amy Hill Bourg	nenis	
			'	,0013	
			Buyer Phone		
			Buyer Email ahill5@lsu.edu	ı	
			Issue Date 02/03/2025		
TITLE:	Motion Capture System				
Addendum 02: Notice is given to all parties that this solicitation is					
amended by the University as stated herein. This Addendum is hereby					
made an official part of this solicitation. See attached addendum.					
To Be Completed By Supplier					
1.	"No Bid" (sign and return this page only).				
2.	My Company does not wish to receive future solicitations for this spend category.				
	Specify your Delivery: To be made within days after receipt of order.				
4.	If applicable, Supplier's Addendum Acknowledgement/Response: As an authorized agent/signatory of the supplier, I/we acknowledge receipt of this Addendum, andsubmit no alterations/clarifications to our original bid.				
	submit superseding revisions/clarifications to our original bid as written herein or attached hereto.				
General Instructions to Suppliers					
	Sealed bids for furnishing the items and/or services specified are hereby solicited, and will be received by LSU Procurement at the "Return Bid To" address stated above, until the specified due date and time.				
2.	Read the entire solicitation, including all terms, conditions and specifications.				
	All bid information and prices must be typed or written in ink. Any corrections, erasures or other forms of alteration to unit price are to be initialed by the supplier.				
	Bid prices are to be quoted FOB LSU/Destination and inclusive of any and all applicable shipping and handling charges unless otherwise specified in the solicitation. Any invoiced delivery charges not quoted and itemized on the LSU purchase order are subject to rejection and non-payment.				
	Payment is to be made within 30 days after receipt of properly executed invoice, or delivery and acceptance, whichever is later.				
	By signing this solicitation, the supplier certifies compliance with all general instructions to suppliers, terms, conditions and specifications; and further certifies that this bid is made without collusion or fraud.				
SUPPLIER NAME MAILIN			AILING ADDRESS		
=					
AUTHORIZED SIGNATURE (TY, STATE ZIP		
PRINTED NAME			HONE #		
TITLE			X #		
E-MAIL I		EDERAL TAX ID #			

LSU RFQ-0000002436

Motion Capture System

Addendum 02

Inquiry 1

Would you be able to clarify the requirement listed as "Real-time multi-user motion tracking coverage (up to three people) with less than 0.02 mm tracking." In particular what is the 0.02 mm tracking referring to?

Response 1

Yes, the 0.02mm refers to the system's accuracy in tracking the markers. Tracking accuracy must be maintained while simultaneously tracking three users wearing motion capture suits in the area covered by the motion capture cameras.

Inquiry 2

In terms of precision, there are different numbers listed throughout the requirements list, including "sub-millimeter" and "0.05" mm. What is the overall system precision requirement?

Response 2

The accuracy of the tracking must be 0.02mm.