



Nicholls State University

Purchasing Office

P. O. Box 2052 – Thibodaux, LA 70310
985.448.4038 – Fax: 985.448.4921

**ADDENDUM NO 1
FOR
NICHOLLS STATE UNIVERSITY
THIBODAUX, LOUISIANA
02/12/2025**

PAGE 1 OF 1 (Plus revised Specification Page following)

Bid SB01896 – Procurement of Hospital Beds and Over Bed Tables

Bid Responses are Due at 3:00 PM on 01/28/2025

This addendum is hereby made a part of the Bidding Documents to the extent as though it was originally included therein.

This addendum, should be acknowledged on the Bid Submission Checklist, Page 13 of 15, and submitted with bid response.

Purpose of Addendum:

1. Provide revised specification page 1 of 1

The original specification page 1 of 1 indicated 20 each for the Overhead Bed Table.

The correct quantity is 43 each per what the bid-form states.

Terry G. Dupre
Director of Purchasing, Property Control
and Support Services Administration

SPECIFICATIONS (Revised 2/12/2025)
Procurement of Hospital Beds and Over Bed Tables
Page 1 of 1

Bid Item 1: Refurbished Hill-Rom P1190 Advanta 2 Med/Surg Hospital Bed. MUST BE THIS BRAND AND MODEL NUMBER TO MATCH EXISTING EQUIPMENT.

Quantity: 43 Each

NOTE: These beds will be used in a classroom/lab setting for instructional purposes only. These beds will not be used in a professional medical/hospital/patient care setting.

- The bed shall be fully refurbished and patient ready
- The bed shall include a brand-new foam mattress
- The bed should include guards on the left and the right
- The bed should include push buttons to adjust the bed
- The bed should be equipped with nurse call integration.]
- The bed Height should be Low Position floor to deck – 15.75", High Position floor to deck 32.5"
- The Sleep Surface should be Width 36", Normal Length 84", Extended Length 88"
- The bed's overall width should be 40", Overall Length normal 84", extended 94"
- The beds' articulation should be:
 - Max head of bed elevation 65°
 - Max knee elevation 36°
 - Foot section inclination -23°
 - Maximum Trendelenburg and Reverse Trend 16°
 - Weigh Limit-safe working load 500 lbs.
- The bed should have a standard non-scale configuration
- The bed shall have a 1-year warranty on parts only
- The Pricing on the Bid form **MUST** include all shipping cost for delivery to Nicholls State University

Bid Item 2: Amico OT-S-TYYN-NU, Overbed Table, NEW, not refurbished. MUST BE THIS BRAND TO MATCH EXISTING EQUIPMENT

Quantity: 43 Each

NOTE: These overbed tables will be used in a classroom/lab setting for instructional purposes only. These overbed tables will not be used in a professional medical/hospital/patient care setting.

- The table should have a single top surface style
- The table should have a standard top profile
- The top should be constructed of a therrmofoil
- The main surface should be flat
- The table finish should be a standard office finish, color to be determined
- The table will not include a basket or vanity
- The base should be a U-Base, Nevada -Color to be determined.
- The table should be shipped assembled.
- The Pricing on the Bid form **MUST** include all shipping cost for delivery to Nicholls State University

