

April 29, 2026

ADDENDUM NO. 1

PROPOSAL FOR FURNISHING, DELIVERING, AND INSTALLING AN HEMATOLOGY ANALYZER (BECKMAN COULTER DxH 690T OR EQUIVALENT) FOR THE UNIVERSITY OF LOUISIANA AT LAFAYETTE NEW IBERIA RESEARCH CENTER LOCATED ON THE AT LAFAYETTE CAMPUS IN NEW IBERIA, LOUISIANA.

Due Monday, May 4, 2026 11:00AM Solicitation No. 26025

The following clauses/alterations shall be made part of the original specifications as though issued at the same time and shall be incorporated integrally therewith.

Item No. 1. –Bid Deadlines:

Bid Due / Bid Opening	Date Changes:
The bid due date was originally scheduled for Thursday, April 30, 2026 at 11:00AM.	The bid due date has been postponed to Monday, May 4, 2026 at 11:00AM.
The bid opening was originally scheduled for Thursday, April 30, 2026 at 2:00PM.	This meeting has been postponed and will now take place on Monday, May 4, 2026 at 2:00PM.
Please note that the Zoom meeting ID and password have not changed and remain the same.	

Item No. 2. – Response to vendor question:

Vendor question (s):	Department response:
<u>Question 1: Terms and Conditions Exceptions:</u> If there are bid terms and conditions the Vendor may not be able to agree to, will the University allow the Vendor to include clarifications or exceptions as part of its bid submission?	Yes, but they will have to be reviewed for acceptance.
<u>Question 2: Insurance Requirements:</u> If there are insurance requirements that the Vendor may not be able to agree to, will the University consider exceptions to insurance terms and conditions?	Yes, we can certainly consider this, and will weigh the risk. Not always can we accommodate, but we can look at the request, weigh the risk and make a decision.
<u>Question 3: Electrical Work Required</u> Is there any electrical work required for installation (beyond plugging into a wall outlet)?	No, unless the instrument requires electrical beyond normal wall outlets (110 or 220), preferably 110.
<u>Question 4: Affixing Equipment:</u> Is affixing the equipment to real property required for installation?	No, this piece of equipment should be a stand alone.

For questions related to bidding these projects, please contact the UL Lafayette Purchasing Department at BidQuestions@louisiana.edu or 337.482.1078.

Business hours are: Mon-Thu 7:30am – 5:00pm CST (lunch 11:45-12:30); Fri 7:30am -12:30pm.

ACKNOWLEDGEMENT: ACKNOWLEDGEMENT: If you have already submitted your bid, and this Addendum creates a need to revise your bid, you must indicate any change(s) below, identify your business name and sign where shown. Revisions shall be submitted/delivered PRIOR to bid due date and time, by email. Bid revisions received after bid due date and time cannot be



Purchasing Office
P.O. Box 40197 • Lafayette, LA 70504-0197
Office: (337) 482-5396
Fax: (337) 482-5059

considered, whereupon the bidder must either honor or withdraw its original bid. If you have already submitted your bid and this addendum does not cause you to revise your bid, acknowledge receipt of this addendum by signing below and returning it to the Purchasing Department prior to bid due date.

Kristi Montet, CPPB
Director for Procurement and Travel
University of Louisiana at Lafayette
Department of Purchasing

Firm Name: _____ **Signature:** _____