

INVITATION TO BID - Addendum 02		LSU	BID DUE DATE AND TIME	
BOARD OF SUPERVISORS OF LOUISIANA STATE UNIVERSITY AND AGRICULTURAL & MECHANICAL COLLEGE			07/30/2026 11:00 AM CT	
SOLICITATION RFQ-0000002867 SUPPLIER # SUPPLIER NAME AND ADDRESS 			RETURN BID TO lsubids@lsu.edu Buyer Amy Hill Bourgeois Buyer Phone Buyer Email ahill5@lsu.edu Issue Date 07/08/2026	
TITLE: Particle Image Velocimetry (PIV) System				
Addendum 02: Notice is given to all parties that this solicitation is amended by the University as stated herein. This Addendum is hereby made an official part of this solicitation. See attached addendum.				
To Be Completed By Supplier				
1. _____ "No Bid" (sign and return this page only). 2. _____ My Company does not wish to receive future solicitations for this spend category. 3. Specify your Delivery: To be made within _____ days after receipt of order. 4. If applicable, Supplier's Addendum Acknowledgement/Response: As an authorized agent/signatory of the supplier, I/we acknowledge receipt of this Addendum, and _____ submit no alterations/clarifications to our original bid. _____ submit superseding revisions/clarifications to our original bid as written herein or attached hereto.				
General Instructions to Suppliers				
1. Sealed bids for furnishing the items and/or services specified are hereby solicited, and will be received by LSU Procurement at the "Return Bid To" address stated above, until the specified due date and time. 2. Read the entire solicitation, including all terms, conditions and specifications. 3. All bid information and prices must be typed or written in ink. Any corrections, erasures or other forms of alteration to unit price are to be initialed by the supplier. 4. Bid prices are to be quoted FOB LSU/Destination and inclusive of any and all applicable shipping and handling charges unless otherwise specified in the solicitation. Any invoiced delivery charges not quoted and itemized on the LSU purchase order are subject to rejection and non-payment. 5. Payment is to be made within 30 days after receipt of properly executed invoice, or delivery and acceptance, whichever is later. 6. By signing this solicitation, the supplier certifies compliance with all general instructions to suppliers, terms, conditions and specifications; and further certifies that this bid is made without collusion or fraud.				
SUPPLIER NAME			MAILING ADDRESS	
AUTHORIZED SIGNATURE			CITY, STATE ZIP	
PRINTED NAME			PHONE #	
TITLE			FAX #	
E-MAIL			FEDERAL TAX ID #	

RFQ-0000002867

Particle Image Velocimetry (PIV) System

Addendum 02

Inquiry 1

The most commonly used high repetition rate lasers output about 30 mJ/pulse at repetition rates up to 1 kHz. With this amount of energy and the camera in side scatter, the maximum illuminated field of view estimate with aerosol seeding is 250 x 500 mm. Will you accept that maximum field of view?

Response 1

No, we will not accept a maximum field of view of 250 x 500 mm. The requirement remains 500 x 1000 mm.

Inquiry 2

If the answer to question #1 is, "no," then are you willing to spend close to \$450,000 for the most powerful laser available to have a chance to get to the 500 x 1000 mm field of view that you desire?

Response 2

No, we are not willing to spend close to \$450,000 for a more powerful laser.